

Certificate

**SMP GmbH
Pruefen Validieren Forschen**

hereby certifies the

cleaning validation of Clear Flush Instruments

of the company

**BOSS Instruments LTD
395 Reas Ford Road
Suite 120
Earlysville; VA 22936
USA**

The SMP investigation, with the project no. 12707011707-2, shows that the cleaning effectiveness of the Medisafe SI PCF Washer in combination with the detergent 3 E Zyme is powerful enough to assure good cleaning results of BOSS Clear Flush instruments without the need of manual pre-cleaning:

- All instruments are visibly clean
- All tips of the instruments are clean after the process.
- All tests fulfil the acceptance criteria

Tests followed

**TS/ISO 15883-5
AAMI TIR 30
AAMI TIR 12**



Tübingen, Germany; Sept. 06. 2007

A handwritten signature in black ink, appearing to read 'K. Roth', is written over a horizontal line.

Klaus Roth
Managing Director